

Exposure control location & operator protection (other than portable units or bone densitometers)

Is exposure control permanently mounted behind a protected barrier - so that the operator can remain completely behind that barrier during the entire exposure?

No

Yes

Is this machine used in Dentistry?

OK - Dosimetry not required unless it is likely the expected dose is greater than 10% of the annual limit (12VAC5-481-760)

No

Yes

Is this a Podiatry or Veterinary unit?

Is exposure control permanently located at least 9 feet from the tube housing assembly?

No

Yes

Violation!

Are means provided to allow the operator to be at least 9 feet from the tube during exposures?
(i.e. - stretch cords)

No

Yes

OK

Was this X-ray installed prior to 5/1/99?

No

Yes

Violation!

Dosimetry required for all operators or holders of patients

No

Yes

Violation!

Are means provided to allow the operator to be at least 9 feet from the tube during exposures?
(i.e. - stretch cords)

No

Yes

Violation!

Dosimetry required for all operators or holders of patients